

Intended use


The SpO₂ adapter cables are intended to be used for connecting reusable or disposable SpO₂ probes to compatible monitoring equipment. These accessories are indicated for use by qualified medical personnel only.

Instructions

1. Connect the adapter cable to the compatible monitor or module.
2. Connect SpO₂ probe to the adapter cable.

Cleaning Procedure

To keep your cable free of dust and dirt, clean it with a lint free cloth moistened with either warm water (40°C maximum) and soap, a diluted non-caustic detergent or one of the approved cleaning agents listed below. Never immerse or soak the cables and connectors in liquid. Do not allow the cleaning agent to remain on the cable – remove it immediately by wiping with a cloth dampened with water.

 If you see signs of deterioration or damage, replace the cable. Do not use it for further patient monitoring.

Recommended Cleaning Agents

- ◆ Mild soap and water solution
- ◆ U.S. Pharmacopeia (USP) green soap
- ◆ Sodium hypochlorite solution (1:10 dilution of household bleach in water)
- ◆ Phenolic germicidal detergent solutions (1% aqueous solutions)
- ◆ Isopropyl alcohol (70%)

Disinfecting Procedure

We recommend that you disinfect the cable only when necessary as determined by your hospital's policy, to avoid long-term damage to the cable. Disinfect your cable with a lint free cloth moistened with one of the approved disinfecting agents listed below. Never immerse or soak the cables and connectors in liquid. Do not allow the disinfectant to remain on the cable – remove it immediately by wiping with a cloth dampened with water.

We also recommend that you clean the cable first as described under “*Cleaning Procedure*”.

Recommended Disinfecting Substances

- ◆ Alcohol based: Ethanol 70%, Isopropanol 70%, Microzid liquid.
- ◆ Aldehyde based: Cidex®

Caution

- ◆ Do not immerse connector ends or cables in liquid.
- ◆ The compatibility of Solaris products with chemical agents other than those listed is unknown.
- ◆ The effectiveness of any cleaning or disinfecting procedure is subject to the proper preparation of each agent per its manufacturer's instructions.
- ◆ Grossly contaminated products which may contact patient during use should be discarded.

Environmental

Operating temperature: 5°C to 40°C

Storage temperature: -20°C to 55°C

Relative humidity: 5% to 95%

Warranty

Solaris offers a 12-month warranty against manufacturing defects in materials and workmanship from the original date of purchase.

