

Certificate



Certificate no.

CU 72092015 01

License Holder:
Apex Medical Corp.
No. 9, Mni Sheng St.
236 Tu-Cheng City, Taipei County
Taiwan - R.O.C.

Manufacturing Plant:
Apex Medical (Shenzhen) Corp.
No. 3, Building
XiliBaimang Xusheng Ind. Estate
Nanshan, Shenzhen
China

Test report no.: USA-DS 30992647 001

Client Reference: Mr. Alan Chang

Tested to: UL 60601-1:2003 R4.06
CAN/CSA-C22.2 No. 601.1-M90

Certified Product: TENS & EMS

License Fee - Units

Model Designation : AP-101020T, AP-101040T, AP-101010T,
TS-1211, TS-1311, AP-101050T,
AP-103000T, TS-1211 COMBO

7

Rated Voltage : DC 9V battery (unit)
AC 100-240V, 50/60Hz (adapter)
Rated Current : 0.8-0.4A (adapter)
Output Ratings DC : see Constructional Data Form

7

Appendix: 1, 1-2

Licensed Test mark:



Signature

Dipl.-Ing. A. Klinker
QA Certification Officer

Date of Issue
(day/mo/yr)
30/09/2009