

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**



MANUFACTURER:

CONTEC MEDICAL SYSTEMS CO., LTD
No.24 Huanghe West Road Economic & Technical
Development Zone ,Qinhuangdao,Hebei Province,
066004,P.R.China

MEDICAL DEVICE:

Pulse Oximeter, SAT-600

CLASSIFICATION - ANNEX IX:

Class II b, Rule 10

CONFORMITY ASSESSMENT ROUTE: Annex II without chapter 4

WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED
MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF
COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES;
INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC
ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH
DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

NOTIFIED BODY:

TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTR 65, D-80339 M NCHEN, GERMANY

IDENTIFICATION NUMBER:

CE 0123

(EC) CERTIFICATE(S):

G1 13 06 50972 019

EC REP

EUROPEAN REPRESENTATIVE:

Shanghai International Trading Corp. GmbH(Hamburg)
Eiffestrasse 80, 20537 Hamburg Germany

START OF CE-MARKING:

2009-07-23 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION:

SIGNATURE:

President

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonized - EN) standards

No.	Serial Number	Title and Description
1	EN 60601-1: 1990+A1:1993+A2:1995	Medical electrical equipment- Part1: General Requirements for Safety
2	EN 60601-1-2: 2007	Medical electrical equipment-Part 1-2: General requirements for basic safety and essential performance-Collateral standard: Electromagnetic compatibility – Requirements and tests
3	EN 60601-1-4:1996+A1: 1999	Medical electrical equipment- Part 1-4: General requirements for safety - Collateral standard: Programmable electrical medical systems
4	EN 60601-1-6:2010 (IEC 60601-1-6:2010)	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability
5	EN 60601-1-8:2007 (IEC 60601-1-8:2006)	Medical electrical equipment - Part 1-8: General requirements for basic safety essential performance- Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems
6	ISO 80601-2-61:2011	Medical electrical equipment —Part 2-61:Particular requirements for basic safety and essential performance of pulse oximeter equipment
7	EN 60601-1-11:2010 (IEC 60601-1-11:2010)	Medical electrical equipment –Part 1-11: General requirements for basic safety and essential performance –Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
8	EN 62366:2008 (IEC 62366:2007)	Medical devices - Application of usability engineering to medical devices
9	EN 62304:2006 (IEC 62304:2006)	Medical device software - Software life-cycle processes