

**EC Certificate**  
**Directive 93/42/EEC Annex V**  
**Production Quality Assurance**  
**Medical Devices**

**Registration No.:** DD 60095048 0001

**Report No.:** 15067822 001

**Manufacturer:** Hangzhou Sejoy Electronics &  
Instruments Co., Ltd.  
Building 2, No. 202, Zhenzhong Road  
West Lake Economy & Technology Zone  
310030 Hangzhou  
China

**Products:** Digital Thermometers, Infra-red Ear  
Thermometers, Blood Pressure Monitors

Replaces Approval, Registration No.: DD 60025446 0001

**Expiry Date:** 2019-06-18

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

**Effective Date:** 2014-07-04

**Date:** 2014-07-04



**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.