



*Declaration of Conformity*

We

3M Health Care

hereby declare under our sole responsibility  
that the CE marked products to which this declaration relates .

Transpore™ surgical tape,  
Product Numbers:  
1527-0, 1527-1, 1527-2, 1527-3, 1527S-1, 1527S-2,  
527-P (1"), 527-P (2"), 771-1PK, 771-2PK, 779,

is classified,  
per Rule 1 of Annex IX of the Medical Device Directive 93/42/EEC,  
as a Class I Non sterile device  
and

is in accordance with Annex VII of Directive 93/42/EEC  
on the approximation of the laws of the European Union Member States  
concerning medical devices.

This certificate is valid for devices originating from the following sites:

3M Brookings  
601 22nd Ave. South  
Brookings, South Dakota 57006 USA

European Representative Address  
3M Health Care  
D41453 Neuss, Germany

Signature: \_\_\_\_\_

Kathryn W. Foran  
3M Health Care  
Regulatory Affairs and Quality Assurance  
Skin & Wound Care Division

Date: \_\_\_\_\_

01/15/2009